

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 07 2009

Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130122</u>	
Name <u>TIM Tuckwood</u>	ID# <u>3446</u> Date <u>6-25-09</u>
A Agency <u>Alaska State Troopers</u> Phone # <u>895-4800</u>	
Instrument Location <u>Delta Junction Court</u>	
B Alco S/N <u>V301596</u> Target Value <u>.081</u> High Pressure <u>550 PSI</u>	
Alco Test Values <u>.082</u> <u>.080</u>	
1 st Alco 2 nd Alco	
Signature <u>Tim Tuckwood</u> <u>CB</u> 7/14/09	
(OVER)	


I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this Jul day of 20th, 2009.

Carolyn M. Noland
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130122 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130122

JUNE 25, 2009

OPERATOR'S NAME:

TUCKWOOD TIM

OPERATOR'S NUMBER: 3446

SUBJECT'S LAST NAME:

VERIF

SUBJECT'S FIRST NAME/MI :

OF CAL

O.L. #: 12345

DEPT/AGENCY: DJT0

DATE/REPORT: 09076

TEST TYPE: U

ALCO TARGET VALUE: .001

ALCO S/N: 4701596

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 28.82 in

ALCO TARGET .078 13:11

BLANK TEST .000 13:12

INTERNAL STANDARD VERIFIED 13:12

ALCO TO 28.82 in .002 13:12

BLANK TEST .000 13:13

SUBJECT SAMPLE .000 13:13

BLANK TEST .000 13:14

ALCO TO 28.81 in .000 13:14

BLANK TEST .000 13:15

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130122

JUNE 25, 2009

TIME 13:16

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 43c

BAROMETER: 28.81 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRIMER TEST

11#02110)++-1/012745473911(=)20000EFG
NICKLECHROMATOPHANT 1211abordergus jkiano
prater 1100111-0